



SITIIAS
Worldwide Access

防爆合格证

证号：GYB23.1528X

制 造 商 常州市武翔仪表有限公司

(地址：江苏省常州市武进区横林镇崔桥街崔蓉路 76 号)

产 品 名 称 隔爆型热电偶

型 号 规 格 aWRbKc-d4Be

防 爆 标 志 Ex db II C T1...T6 Gb, Ex tb III C T80°C Db

产 品 标 准 GB/T 26786-2011

1.144.111~514,

图 样 编 号 WXGBT20190001~20190031,WXGRC20190001~20190003/
20190007~20190009/20190013~20190015/20190019~20190021

经图样及技术文件的审查和样品检验，确认上述产品符合下列标准：

GB/T 3836.1-2021,GB/T 3836.2-2021,GB/T 3836.31-2021

特颁发此证。

本证书有效期：2023 年 01 月 09 日 至 2028 年 01 月 08 日

备注

1. 安全使用注意事项见使用说明书。
2. 证书编号后缀“X”表明产品具有安全使用特殊条件：产品使用环境温度范围： $-20^{\circ}\text{C}\sim+60^{\circ}\text{C}$ ；涉及产品隔爆接合面的维修须联系产品制造厂商。
3. 型号规格说明：a代表热电偶分度号，可为 S、K、N、E、J 或 T；b代表热电偶材料，可为 P、N、M、E、F 或 G；c代表感温元件，可为空白或 2；d代表连接形式：可为 1-无固定、2-螺纹安装和 4-法兰安装；e代表温套管结构形式：可为空白（不选用分体套管时此处不标注）或温套管对应代码（A、B、C、D、E、F、G、H、J、K1、K2、K3、L1、L2、L3、M1、M2、M3、N、P、Q、R、S、T、U、HT1、HT2、HT3、HT4、HT5、HT6）。
4. 外壳防护等级：IP65。




本证书仅对与认可文件和样品一致的产品有效。

地址：上海市漕宝路103号
邮编：200233

网址：www.nepsi.org.cn
Email: info@nepsi.org.cn

电话：+86 21 64368180
传真：+86 21 64844580



EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert No. GYB23.1528X

Manufacturer Changzhou Wuxiang Instrument Co., Ltd.
(Address: No.76 Cuirong Rd, Cuiqiao Street, Henglin Town, Changzhou City, Jiangsu, China)

Product Flameproof Thermocouple

Model aWRbKc-d4Be

Ex marking Ex db IIC T1...T6 Gb, Ex tb IIIC T80°C Db

Product standard GB/T 26786-2011
1.144.111~514,WXGBT20190001~20190031,
WXGRC20190001~20190003/20190007~20190009/
20190013~20190015/20190019~20190021

Drawing number

The product was found to comply with the following standard(s):

GB/T 3836.1-2021,GB/T 3836.2-2021,GB/T 3836.31-2021

Valid until: 2028.01.08

Remarks

- 1.Conditions for safe use are specified in the instructions.
- 2.Symbol "X" placed after the certification number denotes specific conditions of use: Ambient temperature: -20°C~+60°C; Contact the original manufacturer when repair work relates to the flamepath.
- 3.Model designation: a indicates index No of thermocouple, which can be S, K, N, E, J or T; b indicates material, which can be P, N, M, E, F or C; c indicates sensor, which can be blank or 2; d indicates connection, which can be 1, 2 or 4; e indicates protection tube, (A,B,C,D,E,F,G,H,J,K1,K2,K3,L1,L2,L3,M1,M2,M3,N,P,Q,R,S,T,U,HT1,HT2,HT3, HT4,HT5,HT6) .
4. IP code: IP65.



Approval

Shanghai Inspection and Testing Institute of
Instruments and Automation Systems Co., Ltd.
National Supervision and Inspection Center for
Explosion Protection and Safety of Instrumentation
Date of issue 2023.01.09

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel: +86 21 64368180
Fax: +86 21 64844580